

METHOD STATEMENT



Determinand:

Colour Scan

Matrix:

Leachates, effluents and waste waters

Principle of Method:

The sample is filtered through a 0.45µm filter prior to scanning at selected wavelengths between 400 and 800nm at 50nm intervals. For highly coloured samples above the linear range of the spectrophotometer, colour scan on a ten-fold dilution can be requested.

Sampling and Sample Preparation:

Samples should be stored at room temperature and analysed as soon as possible. If samples have been refrigerated, analysis must not be started until a room temperature is attained.

Samples are stable for 30 days (In-House Data) from sampling.

Interferences:

Finely divided suspended matter interferes in the measurement and must be removed by filtration prior to analysis.

Performance of Method:

| Determinand | LOD | Normal Reporting Level | Range of Application |
|----------------|--------|------------------------|----------------------|
| Colour @ 800nm | 0.0030 | 0.005 | 0.005- 2.000 |
| Colour @ 750nm | 0.0038 | 0.005 | 0.005- 2.000 |
| Colour @ 700nm | 0.0028 | 0.005 | 0.005- 2.000 |
| Colour @ 650nm | 0.0022 | 0.005 | 0.005- 2.000 |
| Colour @ 600nm | 0.0018 | 0.005 | 0.005- 2.000 |
| Colour @ 550nm | 0.0019 | 0.005 | 0.005- 2.000 |
| Colour @ 500nm | 0.0018 | 0.005 | 0.005- 2.000 |
| Colour @ 450nm | 0.0017 | 0.005 | 0.005- 2.000 |
| Colour @ 400nm | 0.0044 | 0.005 | 0.005- 2.000 |

| Determinand | MCERTS Accreditation | Low standard | | High Standard | |
|----------------|----------------------|--------------|-------|---------------|-------|
| | | Mean (Abs) | RSD % | Mean (Abs) | RSD % |
| Colour @ 800nm | ✓ | 0.544 | 0.81 | 1.202 | 0.69 |
| Colour @ 750nm | ✓ | 0.455 | 0.82 | 0.989 | 0.65 |
| Colour @ 700nm | ✓ | 0.604 | 0.71 | 1.347 | 1.08 |
| Colour @ 650nm | ✓ | 0.266 | 0.76 | 0.572 | 1.14 |
| Colour @ 600nm | ✓ | 0.102 | 1.52 | 0.213 | 1.81 |
| Colour @ 550nm | ✓ | 0.209 | 1.11 | 0.435 | 1.40 |
| Colour @ 500nm | ✓ | 0.491 | 0.63 | 0.995 | 0.87 |
| Colour @ 450nm | ✓ | 0.380 | 1.08 | 0.763 | 1.14 |
| Colour @ 400nm | ✓ | 0.731 | 2.28 | 1.484 | 1.68 |

METHOD STATEMENT



20% Low Spike

| Determinand | Strensham WTW Trade Effluent | | Finham STW Final Effluent | | Nuneaton STW Final Effluent | |
|----------------|---------------------------------|--------|------------------------------|--------|--------------------------------|--------|
| | %RSD | % Bias | %RSD | % Bias | %RSD | % Bias |
| Colour @ 800nm | 3.89 | 0.50 | 0.85 | -0.48 | 0.78 | -0.21 |
| Colour @ 750nm | 0.53 | -0.28 | 0.87 | -0.47 | 0.75 | -0.19 |
| Colour @ 700nm | 0.99 | 0.35 | 0.68 | 0.37 | 0.60 | 0.11 |
| Colour @ 650nm | 1.01 | 0.62 | 0.81 | 0.24 | 0.69 | -0.13 |
| Colour @ 600nm | 1.88 | 1.23 | 1.57 | -0.06 | 1.48 | -0.71 |
| Colour @ 550nm | 1.65 | 0.34 | 1.28 | -0.35 | 1.13 | -0.02 |
| Colour @ 500nm | 0.99 | 0.18 | 0.62 | -0.14 | 0.57 | 0.17 |
| Colour @ 450nm | 1.10 | -0.15 | 1.03 | -0.78 | 0.99 | -0.41 |
| Colour @ 400nm | 1.83 | -0.85 | 1.51 | -1.42 | 1.50 | -1.29 |

80% High Spike

| Determinand | Strensham WTW Trade Effluent | | Finham STW Final Effluent | | Nuneaton STW Final Effluent | |
|----------------|---------------------------------|--------|------------------------------|--------|--------------------------------|--------|
| | %RSD | % Bias | %RSD | % Bias | %RSD | % Bias |
| Colour @ 800nm | 0.98 | 0.31 | 0.63 | 0.45 | 0.56 | 0.15 |
| Colour @ 750nm | 0.96 | 0.26 | 0.62 | 0.38 | 0.52 | 0.09 |
| Colour @ 700nm | 0.74 | 0.35 | 0.71 | 0.78 | 0.99 | 0.30 |
| Colour @ 650nm | 0.77 | 0.18 | 0.82 | 0.38 | 1.09 | -0.01 |
| Colour @ 600nm | 1.09 | 0.16 | 1.08 | -0.16 | 1.63 | -0.23 |
| Colour @ 550nm | 0.93 | 0.34 | 0.86 | 0.11 | 1.23 | 0.34 |
| Colour @ 500nm | 0.57 | 0.56 | 0.52 | 0.56 | 0.82 | 0.62 |
| Colour @ 450nm | 0.96 | 0.37 | 0.98 | 0.40 | 1.24 | 0.56 |
| Colour @ 400nm | 1.38 | 0.46 | 1.26 | 0.77 | 1.45 | 0.68 |

Uncertainty of Measurement:

The reported uncertainty is an expanded uncertainty calculated using a coverage factor of 2, which gives a level of confidence of approximately 95%.

| Determinand | Uncertainty of Measurement % |
|----------------|------------------------------|
| Colour @ 800nm | 3.39 |
| Colour @ 750nm | 3.31 |
| Colour @ 700nm | 3.35 |
| Colour @ 650nm | 3.48 |
| Colour @ 600nm | 4.93 |
| Colour @ 550nm | 3.92 |
| Colour @ 500nm | 3.37 |
| Colour @ 450nm | 3.79 |
| Colour @ 400nm | 4.46 |

References:

Shimadzu 1601 Operating Manual