

METHOD STATEMENT

Determinand:

Transmission

Matrix:

Waste waters

Principle of Method:

The % T of an unfiltered sample is read using a UV Spectrophotometer set at a wavelength of 254nm.

Sampling and Sample Preparation:

Bring sample to ambient temperature prior to analysis.

Interferences:

None.

Performance of Method:

Performance testing has not been carried out for this method.

References:

Shimadzu 1601 Operating Manual

