



Method Summary

Determination of SVOCs in Aqueous Matrices by GC-MS

Scope and Range

This method describes a procedure for the extraction, detection, identification and quantification of 55 semi-volatile organic compounds, and the semi quantification of other unknown tentatively identified compounds (TIC's). This method holds UKAS accreditation ISO17025 for all compounds except for Phenol, 4-chloroaniline, hexachlorocyclopentadiene, pentachlorophenol and 4-nitrophenol.

The calibration range for this method is LOD to 125µg/l. For limits of detection, see Table 1.

Principle

Preparation and Extraction

A known volume of sample is placed into a separating funnel along with a surrogate solution. The sample is adjusted to pH 7 if required, DCM is added and the sample is shaken with the solvent then removed to a 60ml vial. The sample is then acidified, more DCM added and shaken again. This solvent is then added to the same 60ml vial. The extract is then filtered and concentrated, internal standard added and presented for analysis by GC-MS.

Analysis

Samples are analysed by a gas chromatographic system equipped with an Agilent 6890 Gas Chromatograph along with an Agilent 7683 series auto sampler and an Agilent 5975 inert source mass selective detector.

Interferences

Any compound with similar volatility may co-elute with target analytes and hinder their identification.

Table 1 - Components with Limits of Detection.

COMPOUND	LIMIT OF DETECTION
Phenol	1 µg/l
Bis(2-chloroethyl)ether	1 µg/l
2-Chlorophenol	1 µg/l
1,3-Dichlorobenzene	1 µg/l
1,4-Dichlorobenzene	1 µg/l
1,2-Dichlorobenzene	1 µg/l
2-Methylphenol	1 µg/l
N-nitrosodi-n-propylamine	1 µg/l
Hexachloroethane	1 µg/l
4-Methylphenol	1 µg/l
Nitrobenzene	1 µg/l
Isophorone	1 µg/l
2-Nitrophenol	1 µg/l
2,4-Dimethylphenol	1 µg/l
Bis(2-chloroethoxy)methane	1 µg/l
2,4-Dichlorophenol	1 µg/l
1,2,4-Trichlorobenzene	1 µg/l
Naphthalene	1 µg/l



Method Summary

Determination of SVOCs in Aqueous Matrices by GC-MS

COMPOUND	LIMIT OF DETECTION
4-Chloroaniline	1 µg/l
Hexachlorobutadiene	1 µg/l
4-Chloro-3-methylphenol	1 µg/l
2-Methylnaphthalene	1 µg/l
Hexachlorocyclopentadiene	1 µg/l
2,4,6-Trichlorophenol	1 µg/l
2,4,5-Trichlorophenol	1 µg/l
2-Chloronaphthalene	1 µg/l
2-Nitroaniline	1 µg/l
Dimethyl phthalate	1 µg/l
Acenaphthylene	1 µg/l
2,6-Dinitrotoluene	1 µg/l
3-nitroaniline	1 µg/l
Acenaphthene	1 µg/l
Dibenzofuran	1 µg/l
4-Nitrophenol	1 µg/l
2,4-Dinitrotoluene	1 µg/l
Fluorene	1 µg/l
Diethyl phthalate	1 µg/l
4-Chlorophenylphenylether	1 µg/l
4-Nitroaniline	1 µg/l
Azobenzene	1 µg/l
4-Bromophenylphenylether	1 µg/l
Hexachlorobenzene	1 µg/l
Pentachlorophenol	1 µg/l
Phenanthrene	1 µg/l
Anthracene	1 µg/l
Carbazole	1 µg/l
Di-n-butylphthalate	1 µg/l
Fluoranthene	1 µg/l
Pyrene	1 µg/l
Butylbenzylphthalate	1 µg/l
Benz(a)anthracene	1 µg/l
Chrysene	1 µg/l
Bis(2-ethylhexyl)phthalate	2 µg/l
Di-n-octylphthalate	5 µg/l
Benzo(b)fluoranthene	1 µg/l
Benzo(k)fluoranthene	1 µg/l
Benzo(a)pyrene	1 µg/l
Indeno(1,2,3-cd)pyrene	1 µg/l
Dibenzo(a, h)anthracene	1 µg/l
Benzo(ghi)perylene	1 µg/l