

Method Number: TM422

Updated: 11/07/2022

Issue Number: 04

Page 1 of 1



Method Summary

Determination of Cypermethrin in Liquids by LCMSMS

Scope and Range

This method is used to detect cypermethrin in water samples

The reporting limit for this method is 0.015ng/L

This method is currently accredited to ISO 17025 for ground, surface and saline water

References

none

Principle

Extraction:

Isotope labelled analogues are added prior to sample extraction and carried through the entire extraction procedure and used as internal standards (IS).

A representative portion of the water sample undergoes liquid-liquid extraction and analysis performed on a LC MS MS

Interferences

Interferences co-extracted from the sample will vary considerably from source to source. Resolution via LC and ionisation and detection by ESI-MS-MS reduces the likelihood of this occurring.