

METHOD STATEMENT

Determinand:

Determination of appearance and specific gravity

Matrix:

Sample Type: Treatment chemical samples

Principle of Method:

Some clients send some of their water treatment chemicals for the laboratory to analyse as a check for purity.

Interferences:

None

Performance of Method:

Range of Application:

1.50 to 1.60 g/ml

Limit of Detection, Recoveries of Compounds and Calibration Bias:

Not been determined for this method

Uncertainty of measurement:

Appearance:	N/A
SG:	± 0.001

References:

None

