



## Appendix A to COM 120/19 dated December 16th 2019

Analysis Site	Det Code	Det Description	Unit	Original LoD Reporting Limit	New LoQ Reporting Limit
Otterbourne	60A	Colour	mg/l Pt/Co	0.9	2.8
Otterbourne	60B	Conductivity	µS/cm	1.1	2.2
Otterbourne	60C	pH	pH units	N/A	0
Otterbourne	60D	Turbidity	NTU	0.05	0.16
Wakefield	A00	Propamocarb	µg/l	0.002	0.005
Wakefield	A0A	Hydrogen ion (pH)	pH units	N/A	0
Wakefield	A0B	Conductivity	µS/cm	2.2	8.5
Wakefield	A0C	Turbidity	NTU	0.08	0.15
Wakefield	A0D	Colour	mg/l Pt/Co	1	2.9
Wakefield	A0J	Ammonium ammonia+ammonium ion	mg/l	0.004	0.008
Wakefield	A0K	Nitrite as NO2	mg/l	0.002	0.004
Wakefield	A0L	Nitrate as NO3	mg/l	0.3	0.59
Wakefield	A0M	Orthophosphate as P	mg/l	0.046	0.091
Wakefield	A0P	Chloride as Cl	mg/l	0.6	1.1
Wakefield	A0Q	Aluminium, Filtrate as Al	µg/l	7.22	14.4
Wakefield	A0R	Aluminium, Total as Al	µg/l	7.22	14.4
Wakefield	A0S	Calcium Hardness as CaCO3	mg/l	0.825	1.65
Wakefield	A0T	Calcium, Total as Ca	mg/l	0.33	0.66
Wakefield	A0U	Iron, Filtrate as Fe	µg/l	2.62	5.23
Wakefield	A0V	Iron, Total as Fe	µg/l	2.62	5.23
Wakefield	A0W	Magnesium, Total as Mg	mg/l	0.07	0.13
Wakefield	A0X	Manganese, Filtrate as Mn	µg/l	0.43	0.85
Wakefield	A0Y	Manganese, Total as Mn	µg/l	0.43	0.85
Wakefield	A0Z	Potassium, Total as K	mg/l	0.19	0.37
Wakefield	A16	Sodium, Total as Na	mg/l	0.8	1.59
Wakefield	A18	Antimony, Total as Sb	µg/l	0.04	0.07
Wakefield	A19	Arsenic, Total as As	µg/l	0.08	0.15
Wakefield	A1B	Cadmium, Total as Cd	µg/l	0.013	0.03
Wakefield	A1C	Chromium, Total as Cr	µg/l	0.09	0.18
Wakefield	A1D	Copper, Total as Cu RDT	µg/l	0.2	0.36
Wakefield	A1E	Lead, Total as Pb RDT	µg/l	0.02	0.04
Wakefield	A1F	Nickel, Total as Ni	µg/l	0.09	0.18
Wakefield	A1G	Selenium, Total as Se	µg/l	0.16	0.32
Wakefield	A1H	Silver, Total as Ag	µg/l	0.36	0.72
Wakefield	A1J	Strontium, Total as Sr	µg/l	0.23	0.46
Wakefield	A1K	Zinc, Total as Zn RDT	µg/l	0.68	1.36
Wakefield	A1L	Bromate as BrO3	mg/l	0.0001	0.0003
Wakefield	A1M	Bromide as Br	mg/l	0.003	0.0087
Wakefield	A1N	Chlorate as ClO3	mg/l	0.002	0.0195
Wakefield	A1P	Chlorite as ClO2	mg/l	0.002	0.005
Wakefield	A1Q	Cyanide, Total as CN	µg/l	0.7	1.35
Wakefield	A1R	Fluoride, Total as F	µg/l	20	40
Wakefield	A1S	Mercury, Total as Hg	µg/l	0.03	0.06
Wakefield	A1T	Solids, Dissolved 180 Deg C	mg/l	21.5	43
Wakefield	A1U	Suspended Solids	mg/l	1.09	2.17
Wakefield	A1W	Radioactivity, gross alpha	Bq/l	N/A	0.0157
Wakefield	A1X	Radioactivity, gross beta	Bq/l	N/A	0.009
Wakefield	A1Y	GCMS Full Scan	text	N/A	0
Wakefield	A1Z	Phenol	µg/l	0.037	0.121
Wakefield	A20	2-Chlorophenol	µg/l	0.025	0.083
Wakefield	A21	3-Chlorophenol	µg/l	0.009	0.027
Wakefield	A22	4-Chlorophenol	µg/l	0.01	0.033
Wakefield	A23	2-Chloro,4-methylphenol	µg/l	0.005	0.016

Wakefield	A26	3-Chloro,4-methylphenol	µg/l	0.008	0.024
Wakefield	A27	3-Methylphenol	µg/l	0.013	0.041
Wakefield	A28	2-Methylphenol	µg/l	0.01	0.033
Wakefield	A29	4-Methylphenol	µg/l	0.011	0.035
Wakefield	A2A	2,4-Dichlorophenol	µg/l	0.012	0.039
Wakefield	A2B	2,6-Dichlorophenol	µg/l	0.011	0.03
Wakefield	A2C	2,4,5-Trichlorophenol	µg/l	0.006	0.019
Wakefield	A2D	2,4,6-Trichlorophenol	µg/l	0.007	0.046
Wakefield	A2E	2,3-Dimethylphenol	µg/l	0.005	0.041
Wakefield	A2F	2,5-Dimethylphenol	µg/l	0.005	0.025
Wakefield	A2G	2,6-Dimethylphenol	µg/l	0.009	0.116
Wakefield	A2H	3,4-Dimethylphenol	µg/l	0.006	0.04
Wakefield	A2J	3,5-Dimethylphenol	µg/l	0.005	0.03
Wakefield	A2K	2,4-Dimethylphenol	µg/l	0.005	0.034
Wakefield	A2L	2,3-Dichlorophenol	µg/l	0.004	0.082
Wakefield	A2W	Chlorodibromomethane	µg/l	0.08	0.252
Wakefield	A2X	Bromodichloromethane	µg/l	0.07	0.22
Wakefield	A2Y	Tetrachloroethene	µg/l	0.12	0.37
Wakefield	A2Z	Tetrachloromethane	µg/l	0.04	0.13
Wakefield	A30	Bromoform	µg/l	0.09	0.3
Wakefield	A31	1,1,1-Trichloroethane	µg/l	0.1	0.33
Wakefield	A32	Trichloroethene	µg/l	0.1	0.33
Wakefield	A33	Chloroform	µg/l	0.1	0.31
Wakefield	A34	1,2-Dichloroethane	µg/l	0.13	0.16
Wakefield	A35	Benzene	µg/l	0.03	0.09
Wakefield	A36	Benzo(b)fluoranthene	ng/l	0.52	0.39
Wakefield	A37	Benzo(k)fluoranthene	ng/l	0.47	0.53
Wakefield	A38	Benzo(ghi)perylene	ng/l	0.4	0.43
Wakefield	A39	Benzo(a)pyrene	ng/l	0.22	0.57
Wakefield	A3A	Fluoranthene	ng/l	0.84	0.88
Wakefield	A3B	Indeno(1,2,3cd)pyrene	ng/l	0.48	0.79
Wakefield	A3C	Total PAHs 4 Constituents	µg/l	N/A	0
Wakefield	A49	Metaldehyde	µg/l	0.002	0.005
Wakefield	A4A	Atrazine	µg/l	0.003	0.009
Wakefield	A4B	Cyanazine	µg/l	0.003	0.007
Wakefield	A4C	Simazine	µg/l	0.003	0.008
Wakefield	A4D	Trietazine	µg/l	0.003	0.009
Wakefield	A4E	Carbetamide	µg/l	0.003	0.008
Wakefield	A4F	Chlortoluron	µg/l	0.003	0.008
Wakefield	A4G	Diuron	µg/l	0.005	0.015
Wakefield	A4H	Isoproturon	µg/l	0.003	0.007
Wakefield	A4J	Linuron	µg/l	0.004	0.011
Wakefield	A4K	Monuron	µg/l	0.003	0.008
Wakefield	A4L	Chlorpropham	µg/l	0.004	0.01
Wakefield	A4N	Chlorpyrifos Ethyl	µg/l	0.005	0.014
Wakefield	A4P	Propham	µg/l	0.003	0.009
Wakefield	A4Q	Prosulfocarb	µg/l	0.003	0.008
Wakefield	A4R	Clomazone	µg/l	0.003	0.008
Wakefield	A4S	Difenconazole	µg/l	0.002	0.007
Wakefield	A4T	Flufenacet	µg/l	0.003	0.008
Wakefield	A4U	Flurtamone	µg/l	0.003	0.008
Wakefield	A4X	Quinmerac	µg/l	0.003	0.009
Wakefield	A4Y	Oxadixyl	µg/l	0.004	0.012
Wakefield	A51	Epoxiconazole	µg/l	0.002	0.007
Wakefield	A52	Cyproconazole	µg/l	0.002	0.006
Wavefield	A53	Flusilazole	µg/l	0.003	0.008
Wakefield	A55	Chlorpyrifos Methyl	µg/l	0.004	0.012
Wakefield	A57	Chloridazon	µg/l	0.003	0.007
Wakefield	A6P	IC Screen	Text	N/A	0
Wakefield	A6S	VOC Screen	Text	N/A	0
Wakefield	A6U	UV scan 230-400nm	Text	N/A	0
Wakefield	A6Z	Chlorine, Free	mg/l	0.02	0.03

Wakefield	A70	Chlorine, Total as CL2	mg/l	0.03	0.06
Wakefield	A71	Oil type (qualitative)	Text	N/A	0
Wakefield	ABT	Transmittance @ 254nm	%	N/A	0
Wakefield	ADF	GC MS Scan	Text	N/A	0
Wakefield	ADX	Silica, Total as SiO2	mg/l	0.09	0.15
Wakefield	ADY	Silicate, Total as SiO3	mg/l	0.11	0.19
Wakefield	AE1	Chlorthal dimethyl	µg/l	0.003	0.006
Wakefield	AEK	Ethylbenzene	µg/l	0.03	0.05
Wakefield	AEL	m&p-Xylene	µg/l	0.05	0.14
Wakefield	AEM	o-Xylene	µg/l	0.04	0.13
Wakefield	AEN	MTBE	µg/l	0.29	0.08
Wakefield	AEX	Bromacil	µg/l	0.006	0.02
Wakefield	AEY	Flutriafol	µg/l	0.003	0.008
Wakefield	AEZ	Metazachlor	µg/l	0.003	0.007
Wakefield	AF0	Pendimethalin	µg/l	0.003	0.01
Wakefield	AF1	Propachlor	µg/l	0.003	0.006
Wakefield	AF2	Propiconazole	µg/l	0.002	0.006
Wakefield	AF3	Propyzamide	µg/l	0.004	0.011
Wakefield	AF4	Chlorbufam	µg/l	0.006	0.018
Wakefield	AG2	Toluene	µg/l	0.19	0.63
Wakefield	AG3	Tert-amyl-methyl-ether	µg/l	0.37	0.44
Wakefield	AG4	1,2-Dibromoethane	µg/l	0.42	0.3
Wakefield	AGV	Pentachlorobenzene	µg/l	0.003	0.005
Wakefield	AH1	Fluxapyroxad	µg/l	0.002	0.006
Wakefield	AH3	Metsulfuron-methyl	µg/l	0.003	0.008
Wakefield	AH4	Picloram	µg/l	0.006	0.019
Wakefield	AH7	Terbuthylazine	µg/l	0.001	0.003
Wakefield	AH8	Vinyl Chloride	µg/l	0.056	0.123
Wakefield	AH9	Metaldehyde	µg/l	0.003	0.01
Wakefield	AHA	2,3,6-Trichlorobenzoic Acid	µg/l	0.003	0.01
Wakefield	AHB	2,4,5-T	µg/l	0.005	0.016
Wakefield	AHC	2,4,5-TP (Fenoprop)	µg/l	0.006	0.018
Wakefield	AHD	2,4-D	µg/l	0.004	0.012
Wakefield	AHE	2,4-DB	µg/l	0.006	0.018
Wakefield	AHF	2,4-DP (Dichloroprop)	µg/l	0.005	0.017
Wakefield	AHG	Bentazone	µg/l	0.002	0.004
Wakefield	AHH	Bromoxynil	µg/l	0.003	0.009
Wakefield	AHJ	Clopyralid	µg/l	0.005	0.015
Wakefield	AHK	Dicamba	µg/l	0.004	0.012
Wakefield	AHL	Ioxynil	µg/l	0.003	0.009
Wakefield	AHM	MCPA	µg/l	0.004	0.012
Wakefield	AHN	MCPB	µg/l	0.006	0.018
Wakefield	AHP	MCPP (Mecoprop)	µg/l	0.003	0.009
Wakefield	AHQ	Pentachlorophenol	µg/l	0.006	0.02
Wakefield	AHR	Triclopyr	µg/l	0.009	0.028
Wakefield	AHS	Benazolin	µg/l	0.003	0.01
Wakefield	AHT	Fluroxypyr	µg/l	0.003	0.009
Wakefield	AHU	Imazapyr	µg/l	0.003	0.009
Wakefield	AHV	Asulam	µg/l	0.003	0.008
Wakefield	AHW	Microcystin-LR	µg/l	0.01	0.031
Wakefield	AHZ	Ethoprophos	µg/l	0.002	0.006
Wakefield	AJ0	Lenacil	µg/l	0.002	0.006
Wakefield	AJ2	Metribuzin	µg/l	0.001	0.004
Wakefield	AJ3	Spiroxamine	µg/l	0.001	0.002
Wakefield	AJ7	Parathion-Ethyl	µg/l	0.004	0.007
Wakefield	AJA	Silica, reactive as Si	mg/l	0.01	0.03
Wakefield	AJB	Silica, reactive as SiO2	mg/l	0.023	0.046
Wakefield	AJD	Ametryn	µg/l	0.002	0.004
Wakefield	AJE	Azinphos-methyl	µg/l	0.001	0.003
Wakefield	AJF	Carbendazim	µg/l	0.001	0.002
Wakefield	AJG	Chlorfenvinphos	µg/l	0.003	0.01
Wakefield	AJJ	Dichlorvos	µg/l	0.002	0.006

Wakefield	AJK	Dimethoate	µg/l	0.002	0.004
Wakefield	AJM	Ethofumesate	µg/l	0.002	0.005
Wakefield	AJN	Fenpropidin	µg/l	0.001	0.004
Wakefield	AJP	Fenpropimorph	µg/l	0.001	0.004
Wakefield	AJS	Malathion	µg/l	0.002	0.004
Wakefield	AJU	Methabenzthiazuron	µg/l	0.001	0.004
Wakefield	AJV	Methiocarb	µg/l	0.003	0.008
Wakefield	AJW	Mevinphos	µg/l	0.002	0.005
Wakefield	AJY	Pirimicarb	µg/l	0.001	0.003
Wakefield	AJZ	Pirimiphos-methyl	µg/l	0.001	0.004
Wakefield	AK1	Aldrin	µg/l	0.003	0.005
Wakefield	AK4	Alpha-HCH	µg/l	0.003	0.005
Wakefield	AK8	Chlorothalonil	µg/l	0.007	0.014
Wakefield	AK9	cis-Heptachlor Epoxide	µg/l	0.003	0.006
Wakefield	AKA	cis-Permethrin	µg/l	0.003	0.006
Wakefield	AKB	Cyfluthrin	µg/l	0.003	0.006
Wakefield	AKC	Cypermethrin	µg/l	0.003	0.005
Wakefield	AKE	Deltamethrin	µg/l	0.003	0.005
Wakefield	AKF	Dichlobenil	µg/l	0.003	0.005
Wakefield	AKG	Dieldrin	µg/l	0.003	0.005
Wakefield	AKH	Endrin	µg/l	0.003	0.005
Wakefield	AKK	Fenvalerate	µg/l	0.004	0.007
Wakefield	AKL	Gamma-HCH	µg/l	0.003	0.006
Wakefield	AKM	Heptachlor	µg/l	0.002	0.004
Wakefield	AKN	Hexachlorobenzene	µg/l	0.002	0.004
Wakefield	AKR	Methoxychlor	µg/l	0.004	0.007
Wakefield	AKS	op-DDE	µg/l	0.003	0.005
Wakefield	AKT	op-DDT	µg/l	0.003	0.005
Wakefield	AKU	op-TDE	µg/l	0.003	0.005
Wakefield	AL3	pp-DDE	µg/l	0.003	0.005
Wakefield	AL4	pp-DDT	µg/l	0.003	0.005
Wakefield	AL5	pp-TDE	µg/l	0.003	0.005
Wakefield	AL6	Tecnazene	µg/l	0.004	0.007
Wakefield	AL7	trans-Heptachlor Epoxide	µg/l	0.002	0.005
Wakefield	AL8	trans-Permethrin	µg/l	0.003	0.005
Wavefield	AL9	Trifluralin	µg/l	0.002	0.004
Wakefield	ALA	Prometryn	µg/l	0.001	0.003
Wakefield	ALB	Propazine	µg/l	0.001	0.003
Wakefield	ALC	Propetamphos	µg/l	0.003	0.01
Wakefield	ALD	Terbutryn	µg/l	0.001	0.002
Wakefield	ALF	Triazophos	µg/l	0.002	0.004
Wakefield	ALH	Total Solids @ 105C	mg/l	21.5	43
Wakefield	ALJ	Sodium, Filtrate as Na	mg/l	0.8	1.59
Wakefield	ALK	Chromium, Filtrate as Cr	µg/l	0.09	0.18
Wakefield	ALL	Nickel, Filtrate as Ni	µg/l	0.09	0.18
Wakefield	ALM	Silver, Filtrate as Ag	µg/l	0.36	0.72
Wakefield	ALN	Zinc, Filtrate as Zn	µg/l	0.68	1.36
Wakefield	ALP	Arsenic, Filtrate as As	µg/l	0.08	0.15
Wakefield	ALQ	Cadmium, Filtrate as Cd	µg/l	0.02	0.03
Wakefield	ALR	Phenols, Total Low Level	µg/l	N/A	0
Wakefield	ALS	Hardness, Total as CaCO3	mg/l	1.12	2.19
Wakefield	ALT	Chlorine, Combined	mg/l	0.03	0.06
Wakefield	ALU	Magnesium Hardness as CaCO3	mg/l	0.287	0.533
Wakefield	ALV	Total Inorganic Phosphate as P	mg/l	0.027	0.091
Wakefield	ALW	Total Inorganic Phosphate as P	µg/l	27	91
Wakefield	ALX	Xylene, Total	µg/l	N/A	0
Wakefield	ALZ	Alpha Emitter Scan	N/A	N/A	NULL
Wakefield	AMO	Metal Scan	TEXT	N/A	0
Wakefield	AMM	Boron, Total as Borate	µg/l	86.7	173.2
Wakefield	AMS	DiChloroMethane	µg/l	0.58	0.38
Wakefield	AMT	11-DiChloroEthane	µg/l	0.02	0.07
Wakefield	AMU	14-DiChloroBenzene	µg/l	0.26	0.32

Wakefield	AMV	Total Organic Carbon	mg/l	0.35	0.7
Wakefield	AMW	Dissolved Organic Carbon	mg/l	0.35	0.7
Wakefield	AN0	Dalapon	µg/l	0.2	0.6
Wakefield	AN1	MonoChloroAcetic Acid	µg/l	0.2	0.6
Wakefield	AN2	TriBromoAcetic Acid	µg/l	0.4	1.1
Wakefield	AN3	DiBromoChloroAcetic Acid	µg/l	0.3	0.8
Wakefield	AN4	BromoDiChloroAcetic Acid	µg/l	0.2	0.5
Wakefield	AN5	DiBromoAcetic Acid	µg/l	0.1	0.4
Wakefield	AN6	TriChloroAcetic Acid	µg/l	0.1	0.3
Wakefield	AN7	BromoChloroAcetic Acid	µg/l	0.2	0.4
Wakefield	AN8	DiChloroAcetic Acid	µg/l	0.2	0.6
Wakefield	AN9	MonoBromoAcetic Acid	µg/l	0.2	0.5
Wakefield	ARW	Aluminium, Total as Al, Unflus	µg/l	7.22	14.4
Wakefield	ARX	Iron, Total as Fe, Unflushed	µg/l	2.62	5.23
Wakefield	ARY	Atrazine Desethyl	µg/l	0.001	0.003
Wakefield	ARZ	Atrazine Desisopropyl	µg/l	0.001	0.003
Wakefield	AS5	Acrylamide	µg/l	0.003	0.01
Wakefield	ASF	PAH Total µg/l	µg/l	N/A	0
Wakefield	ATD	Lead, filtrate as Pb	µg/l	0.02	0.04
Wakefield	ATG	Pb Tot After 2 Min Flush	µg/l	0.02	0.04
Wakefield	ATJ	Pb Sol After 2 Min Flush	µg/l	0.02	0.04
Wakefield	AUB	Lithium, Total as Li	µg/l	1.17	2.34
Wakefield	AUC	Beryllium, Total as Be	µg/l	0.1	0.19
Wakefield	AUD	Titanium, Total as Ti	µg/l	0.43	0.86
Wakefield	AUE	Vanadium, Total as V	µg/l	0.06	0.11
Wakefield	AUF	Cobalt, Total as Co	µg/l	0.04	0.08
Wakefield	AUG	Molybdenum, Total as Mo	µg/l	0.38	0.76
Wakefield	AUH	Tin, Total as Sn	µg/l	0.66	1.31
Wakefield	AUJ	Barium, Total as Ba	µg/l	0.21	0.42
Wakefield	AUK	Thallium, Total as Tl	µg/l	0.08	0.15
Wakefield	AUL	Uranium, Total as U	µg/l	0.02	0.03
Wakefield	AVD	Total Oxidised Nitrogen as NO3	mg/l	0.29	0.59
Wakefield	AVF	Strontium, Total as Sr	mg/l	0.0003	0.0005
Wakefield	AVG	Orthophosphate as P	mg/l	0.046	0.091
Wakefield	AVK	THM, Total	µg/l	N/A	0
Wakefield	AVL	Tetra and Trichloroethene	µg/l	N/A	0
Wakefield	AVM	Nitrogen Formula	NULL	N/A	NULL
Wakefield	AVS	Fluoride, Total as F	mg/l	0.02	0.04
Wakefield	AVU	Benzo(b)fluoranthene	µg/l	0.00052	0.00039
Wakefield	AVV	Benzo(k)fluoranthene	µg/l	0.00047	0.00053
Wakefield	AVW	Benzo(ghi)perylene	µg/l	0.0004	0.00043
Wakefield	AVX	Benzo(a)pyrene	µg/l	0.00022	0.00057
Wakefield	AVY	Fluoranthene	µg/l	0.00084	0.00088
Wakefield	AVZ	Indeno(1,2,3cd)pyrene	µg/l	0.00048	0.00079
Wakefield	AW0	Hardness, Total as Ca	mg/l	0.446	0.876
Wakefield	AW1	Copper, Total as Cu RDT	mg/l	0.18	0.36
Wakefield	AW3	Boron, Total as B	mg/l	0.015	0.032
Wakefield	AW4	Copper, Filtered as Cu	µg/l	0.18	0.36
Wakefield	AW5	Molybdenum, Total as Mo	mg/l	0.0004	0.0008
Wakefield	AW6	Orthophosphate as P	µg/l	46	91
Wakefield	AW9	Copper, Total as Cu	µg/l	0.18	0.36
Wakefield	AWA	Lead, Total as Pb	µg/l	0.02	0.04
Wakefield	AWB	Zinc, Total as Zn	µg/l	0.68	1.36
Wakefield	AWC	Copper, Total as Cu	mg/l	0.0002	0.0004
Wakefield	AWK	Light Abs 254nm	Abs/m	N/A	0
Wakefield	AWL	Orthophosphate, filtered as P	mg/l	0.046	0.091
Wakefield	AWS	4,2 and 2,5 Chloromethylphenol	µg/l	0.008	0.026
Wakefield	AX0	Turbidity Ex - works	NTU	0.07	0.15
Wakefield	AX1	Colour, unfiltered	mg/l Pt/Co	1	2.9
Wakefield	AX3	Solids, Tot Dissolved 105DegC	mg/l	21.5	43
Wakefield	AX4	Solids, Tot Dissolved 260 DegC	mg/l	21.5	43
Wakefield	AX5	Nitrite/Nitrate Ratio	mg/l	N/A	NULL

Wakefield	AX6	Cyanide, Free as CN	µg/l	0.52	1.05
Wakefield	AX8	Bicarbonate as CaCO3	mg/l	10.4	20.8
Wakefield	AX9	Light Abs 254nm	Abs/cm	N/A	0
Wakefield	AXD	Nickel, Total as Ni, RDT	µg/l	0.09	0.18
Wakefield	AXE	Phosphorus, Total as P	µg/l	11.41	22.81
Wakefield	AXF	Phosphorus, Filtered as P	µg/l	11.41	22.81
Wakefield	AXG	Sulphur, Total as SO4	mg/l	0.34	0.68
Wakefield	AXH	Silicon, Total as Si	mg/l	0.03	0.07
Wakefield	AXW	EPTC	µg/l	0.004	0.013
Wakefield	AXZ	Alkalinity as CaCO3	mg/l	10.4	20.8
Wakefield	AY0	Permanganate Index	mg/l	0.29	0.59
Wakefield	AY1	Alkalinity as HCO3	mg/l	12.7	25.3
Wakefield	AY2	Bromate as BrO3	µg/l	0.0753	0.26
Wakefield	AY3	Bromide as Br	µg/l	1.7	8.7
Wakefield	AY4	Chlorate as ClO3	µg/l	5.84	19.5
Wakefield	AY5	Chlorite as ClO2	µg/l	1.42	4.8
Wakefield	AY6	Sulphide	µg/l	10	20
Wakefield	AY8	Total Indicative Dose	Text	N/A	NULL
Wakefield	AY9	Calcium, Filtrate as Ca	mg/l	0.31	0.66
Wakefield	AYA	Boron, Filtrate as B	µg/l	14.9	31.9
Wakefield	AYC	Permethrin, Total	µg/l	N/A	0
Wakefield	AYJ	Boron, Total as B	µg/l	15.93	31.85
Wakefield	AYK	Boron, Filtrate as B	µg/l	15.93	31.85
Wakefield	AYM	Magnesium, Filtrate as Mg	mg/l	0.07	0.13
Wakefield	AYP	Oil Identification	Text	N/A	0
Wakefield	AZP	Boscalid	µg/l	0.004	0.011
Wakefield	AZQ	Metoxuron	µg/l	0.002	0.005
Wakefield	AZR	Tebuconazole	µg/l	0.002	0.007
Wakefield	AZW	Tin, Filtrate as Sn	µg/l	0.66	1.31
Wakefield	D1L	Diazinon	µg/l	0.003	0.006
Wakefield	D1M	Diffufenican	µg/l	0.004	0.007
Wakefield	D1N	Kresoxim Methyl	µg/l	0.003	0.005
Wakefield	D1P	Lambda-Cyhalothrin	µg/l	0.003	0.006
Wakefield	D1Q	Triallate	µg/l	0.003	0.005
Wakefield	EK0	Emergency Test Kit, Cr(VI)	mg/l	N/A	0.01
Wakefield	EK1	Emergency Test Kit, Free CN	mg/l	N/A	0.1
Wakefield	EK2	Emergency Test Kit, Formald	mg/l	N/A	0.13
Wakefield	EK3	Emergency Test Kit, Sulphide	mg/l	N/A	0.1
Wakefield	Y2X	Total Heptachlor Epoxide	µg/l	N/A	0
Wakefield	AQ0	PHENOL	µg/l	0.012	0.038
Wakefield	AQ1	2-METHYLPHENOL	µg/l	0.003	0.01
Wakefield	AQ2	3-METHYLPHENOL	µg/l	0.004	0.013
Wakefield	AQ3	4-METHYLPHENOL	µg/l	0.003	0.007
Wakefield	AQ4	2-CHLOROPHENOL	µg/l	0.002	0.007
Wakefield	AQ5	3-CHLOROPHENOL	µg/l	0.002	0.007
Wakefield	AQ6	4-CHLOROPHENOL	µg/l	0.003	0.007
Wakefield	AQ7	2,6-DIMETHYLPHENOL	µg/l	0.002	0.006
Wakefield	AQ8	2,5-DIMETHYLPHENOL	µg/l	0.002	0.006
Wakefield	AQ9	2,4-DIMETHYLPHENOL	µg/l	0.003	0.007
Wakefield	AQA	3,5-DIMETHYLPHENOL	µg/l	0.003	0.008
Wakefield	AQB	2,3-DIMETHYLPHENOL	µg/l	0.002	0.005
Wakefield	AQC	3,4-DIMETHYLPHENOL	µg/l	0.002	0.007
Wakefield	AQD	4-CHLORO-2-METHYLPHENOL + 2-CHLORO-5-I	µg/l	0.005	0.015
Wakefield	AQE	2-CHLORO-4-METHYLPHENOL	µg/l	0.003	0.009
Wakefield	AQF	3-CHLORO-4-METHYLPHENOL	µg/l	0.005	0.016
Wakefield	AQG	2,6-DICHLOROPHENOL	µg/l	0.002	0.007
Wakefield	AQH	2,4-DICHLOROPHENOL	µg/l	0.015	0.049
Wakefield	AQJ	2,3-DICHLOROPHENOL	µg/l	0.003	0.01
Wakefield	AQK	2,4,6-TRICHLOROPHENOL	µg/l	0.007	0.022
Wakefield	AQL	2,4,5-TRICHLOROPHENOL	µg/l	0.003	0.01
Wakefield	D4A	AMETRYN	µg/l	0.001	0.003
Wakefield	D4B	ATRAZINE	µg/l	0.002	0.004

Wakefield	D4C	ATRAZINE-DESETHYL	µg/l	0.002	0.005
Wakefield	D4D	ATRAZINE-DEISOPROPYL	µg/l	0.003	0.007
Wakefield	D4E	AZINPHOS-METHYL	µg/l	0.001	0.003
Wakefield	D4F	AZOXYSTROBIN	µg/l	0.003	0.008
Wakefield	D4G	BIXAFEN	µg/l	0.004	0.014
Wakefield	D4H	BOSCALID	µg/l	0.004	0.011
Wakefield	D4J	BROMACIL	µg/l	0.003	0.009
Wakefield	D4K	CARBENDAZIM	µg/l	0.001	0.003
Wakefield	D4L	CARBETAMIDE	µg/l	0.003	0.008
Wakefield	D4M	CHLORBUFAM	µg/l	0.006	0.02
Wakefield	D4N	CHLORFENVINPHOS	µg/l	0.004	0.012
Wakefield	D4P	CHLORIDAZON	µg/l	0.003	0.008
Wakefield	D4Q	CHLORPROPHAM	µg/l	0.004	0.012
Wakefield	D4R	CHLORPYRIPHOS-ETHYL	µg/l	0.005	0.016
Wakefield	D4S	CHLORPYRIPHOS-METHYL	µg/l	0.006	0.018
Wakefield	D4T	CHLORTOLURON	µg/l	0.002	0.007
Wakefield	D4U	CLOMAZONE	µg/l	0.002	0.005
Wakefield	D4V	CYANAZINE	µg/l	0.002	0.006
Wakefield	D4W	CYPROCONAZOLE	µg/l	0.002	0.007
Wakefield	D4X	DICHLORVOS	µg/l	0.003	0.008
Wakefield	D4Y	DIFENOCONAZOLE	µg/l	0.003	0.009
Wakefield	D4Z	DIMETHOATE	µg/l	0.002	0.007
Wakefield	D50	DIURON	µg/l	0.003	0.009
Wakefield	D51	EPOXICONAZOLE	µg/l	0.002	0.005
Wakefield	D52	EPTC	µg/l	0.003	0.01
Wakefield	D53	ETHOFUMESATE	µg/l	0.005	0.015
Wakefield	D54	ETHOPROPHOS	µg/l	0.003	0.008
Wakefield	D55	FENPROPIDIN	µg/l	0.003	0.007
Wakefield	D56	FENPROPIMORPH	µg/l	0.002	0.004
Wakefield	D57	FLUFENACET	µg/l	0.003	0.008
Wakefield	D58	FLURTAMONE	µg/l	0.003	0.007
Wakefield	D59	FLUSILAZOLE	µg/l	0.002	0.006
Wakefield	D5A	FLUTRIFOL	µg/l	0.002	0.005
Wakefield	D5B	FLUXAPYROXAD	µg/l	0.003	0.009
Wakefield	D5C	ISOPROTURON	µg/l	0.002	0.005
Wakefield	D5D	LENACIL	µg/l	0.002	0.007
Wakefield	D5E	LINURON	µg/l	0.004	0.012
Wakefield	D5F	MALATHION	µg/l	0.003	0.007
Wakefield	D5G	METAZACHLOR	µg/l	0.002	0.006
Wakefield	D5H	METCONAZOLE	µg/l	0.002	0.005
Wakefield	D5J	METHABENZTHIAZURON	µg/l	0.002	0.005
Wakefield	D5K	METHIOCARB	µg/l	0.003	0.008
Wakefield	D5L	METOXURON	µg/l	0.002	0.007
Wakefield	D5M	METRIBUZIN	µg/l	0.002	0.005
Wakefield	D5N	METSULFURON-METHYL	µg/l	0.006	0.019
Wakefield	D5P	MEVINPHOS	µg/l	0.002	0.005
Wakefield	D5Q	MONURON	µg/l	0.002	0.007
Wakefield	D5R	MYCLOBUTANIL	µg/l	0.003	0.008
Wakefield	D5S	OXADIXYL	µg/l	0.004	0.012
Wakefield	D5T	PENDIMETHALIN	µg/l	0.006	0.02
Wakefield	D5U	PIRIMICARB	µg/l	0.002	0.005
Wakefield	D5V	PIRIMIPHOS-METHYL	µg/l	0.002	0.004
Wakefield	D5W	PROCHLORAZ	µg/l	0.002	0.006
Wakefield	D5X	PROMETRYN	µg/l	0.002	0.004
Wakefield	D5Y	PROPACHLOR	µg/l	0.002	0.006
Wakefield	D5Z	PROPAZINE	µg/l	0.001	0.004
Wakefield	D61	PROPETAMPHOS	µg/l	0.005	0.015
Wakefield	D62	PROPHAM	µg/l	0.002	0.007
Wakefield	D63	PROPICONAZOLE	µg/l	0.003	0.007
Wakefield	D64	PROPYZAMIDE	µg/l	0.003	0.009
Wakefield	D65	PROSULFOCARB	µg/l	0.003	0.009
Wakefield	D66	PROTHIOCONAZOLE-DESTHIO	µg/l	0.002	0.005

Wakefield	D67	QUINMERAC	µg/l	0.002	0.006
Wakefield	D68	SIMAZINE	µg/l	0.002	0.004
Wakefield	D69	SPIROXAMINE	µg/l	0.002	0.006
Wakefield	D6A	TEBUCONAZOLE	µg/l	0.002	0.005
Wakefield	D6B	TERBUTHYLAZINE	µg/l	0.001	0.003
Wakefield	D6C	TERBUTRYN	µg/l	0.001	0.003
Wakefield	D6D	TRIAZOPHOS	µg/l	0.002	0.006
Wakefield	D6E	TRIAZINE	µg/l	0.002	0.004
Wakefield	D6F	TRINEXAPAC-ETHYL	µg/l	0.004	0.012
Wakefield	ANP	AMPA	µg/l	0.003	0.007
Wakefield	ANN	GLYPHOSATE	µg/l	0.002	0.005
Wakefield	ANQ	CHLORMEQUAT	µg/l	0.001	0.003
Wakefield	ANR	MEPIQUAT	µg/l	0.002	0.004
Wakefield	ANS	2-BromoPhenol	µg/l	0.002	0.004
Wakefield	ANT	2,6-DiBromoPhenol	µg/l	0.001	0.003
Wakefield	ANU	2,4-DiBromoPhenol	µg/l	0.001	0.003
Wakefield	ANV	246-TriBromoPhenol	µg/l	0.001	0.003
Wakefield	ANW	2-IsoProp-3-Me-Pyr	µg/l	0.001	0.003
Wakefield	ANX	3-ChloroAnisole	µg/l	0.001	0.003
Wakefield	ANY	4-ChloroAnisole	µg/l	0.001	0.003
Wakefield	ANZ	2-ChloroAnisole	µg/l	0.001	0.003
Wakefield	AP5	2-IsoBut-3-Me-Pyr	µg/l	0.001	0.003
Wakefield	AP6	2-MIB	µg/l	0.001	0.003
Wakefield	AP7	246-TriChloroAnisole	µg/l	0.001	0.003
Wakefield	AP8	Geosmin	µg/l	0.001	0.003
Wakefield	AP9	234-TriChloroAnisole	µg/l	0.001	0.003
Wakefield	APA	246-TriBromoAnisole	µg/l	0.001	0.002
Wakefield	APC	Geosmin	ng/l	1	3
Wakefield	APD	2-MIB	ng/l	1	3